

LightCycler™ -Primer Set

Ready-to-use amplification primer mix for RT-PCR using the LightCycler™ Instrument

Human Bcl-10

Kit for 96 reactions

Lot # 191202 Exp.19.12.2003

Note: After Thawing keep on ice!

Store the kit at -20°C

| | | | | |
|---|---|-------------------------------------|---|---|
| 1.Kit Contents | | | | |
| caution | After Thawing keep on ice! | | Sample material | |
| Kit contents | Vial | Label | Sample Preparation | |
| | 1 | Bcl-10 Primer mix Yellow cap | | <p>cDNA reverse transcribed from human RNA</p> <p>Reliable and reproducible results are achieved with 1µg total RNA isolated with the HighPure total RNA Isolation Kit (Roche) reverse transcribed with the 1st Strand cDNA Synthesis Kit (AMV) (Roche).</p> <p>! The resulting cDNA has to be diluted to a final volume of 200-500 µl with PCR-grade water</p> |
| | 2 | Standard Red cap | | |
| | 3 | Standard Stabilizer Green cap | | |
| | 4 | Control cDNA Blue cap | | |
| 5 | H2O, sterile, PCR grade White cap | | | |
| Additional equipment and reagents required | 1 st Strand cDNA Synthesis Kit for RT-PCR (Roche Cat. # 1 483 188) LightCycler™ FastStart Master SybrGreen I (Roche Cat. # 3 003 230) LightCycler™ Instrument (Roche Cat. # 2 011 468) LightCycler™ Primer Set Housekeeping genes (Search GmbH) | | Application | |
| 2. Introduction | | | Assay time | |
| <p>The LightCycler™ -Primer Set allows to perform quantitative RT-PCR using the LightCycler™ instrument. An optimized primer pair has been selected for specific amplification of targets. The amplicon is detected by fluorescence using the double-stranded DNA binding dye Sybr®Green I.</p> | | | <p>Set up the PCR amplification 15 min LightCycler™ PCR run 50 min</p> | |
| | | | Number of tests | |
| | | | Quality Control | |
| | | | Kit storage/stability | |
| | | | Specificity | |
| | | | <p>The Kit is designed for 96 Reactions</p> <p>The LightCycler™ -Primer Set is tested using the LightCycler™ FastStart Master Sybr®Green I according to the protocol described below.</p> <p>The unopened kit is stable at -20°C 12 month from date of manufacture</p> <p>The LightCycler™ -Primer Set "Bcl-10" is specific for the sequence of human Bcl-10 and does not detect genomic Bcl-10 specific sequences if used as directed.</p> | |

3. Procedure

| | | | | | | | |
|--|--|--------------|-------------|--|------------------|-------------------------|-----------|
| Introduction | A fragment of the human Bcl-10 cDNA sequence is amplified and monitored with the dsDNA specific Sybr®Green I dye | | | | | | |
| Additional reagents required | LightCycler™ FastStart Master Sybr®Green I (Cat.# 3 003 230) | | | | | | |
| Thawing the solutions | Thaw the following reagents, mix gently, and store on ice: <table border="0"> <tr> <td>From the ...</td> <td>Thaw the...</td> </tr> <tr> <td>LightCycler™ FastStart Master Sybr®Green I</td> <td>vial 1a/b</td> </tr> <tr> <td>LightCycler™ Primer Set</td> <td>all tubes</td> </tr> </table> <p>It is recommended to define the experimental protocol before preparing the solutions</p> | From the ... | Thaw the... | LightCycler™ FastStart Master Sybr®Green I | vial 1a/b | LightCycler™ Primer Set | all tubes |
| From the ... | Thaw the... | | | | | | |
| LightCycler™ FastStart Master Sybr®Green I | vial 1a/b | | | | | | |
| LightCycler™ Primer Set | all tubes | | | | | | |
| Experimental Protocol | The described protocol consists of four programs. <ul style="list-style-type: none"> • Program 1: Denaturation of the template and activation of the polymerase • Program 2: Amplification of the target • Program 3: Melting curve analysis for product control • Program 4: Cooling the rotor and thermal chamber | | | | | | |

Denaturation

| Parameter | Value |
|------------------------------|-----------|
| Cycles | 1 |
| Type | Regular |
| Temp. Targets | Segment 1 |
| Target Temperature | 95 |
| Incubation time (h:min:s) | 10:00 |
| Temp. Transition Rate (°C/s) | 20 |
| Secondary Target Temp. | 0 |
| Step Size | 0 |
| Step Delay | 0 |
| Aquisition Mode | None |

Amplification

| Parameter | Value | | |
|------------------------------|----------------|-------|--------|
| Cycles | 35 | | |
| Type | Quantification | | |
| Temp. Targets | Seg.1 | Seg.2 | Seg.3 |
| Target Temperature | 95 | 68 | 72 |
| Incubation time (h:min:s) | 10 | 10 | 16 |
| Temp. Transition Rate (°C/s) | 20 | 20 | 20 |
| Secondary Target Temp. | 0 | 58 | 0 |
| Step Size | 0 | 0.5 | 0 |
| Step Delay | 0 | 1 | 0 |
| Aquisition Mode | None | None | Single |
| Gains | F1 = 5 | | |

Melting Curve Analysis

| Parameter | Value | | |
|------------------------------|---------------|--------|-------|
| Cycles | 1 | | |
| Type | Melting Curve | | |
| Temp. Targets | Seg.1 | Seg. 2 | Seg.3 |
| Target Temperature | 95 | 58 | 95 |
| Incubation time (h:min:s) | 0 | 10 | 0 |
| Temp. Transition Rate (°C/s) | 20 | 20 | 0.1 |
| Secondary Target Temp. | 0 | 0 | 0 |
| Step Size | 0 | 0 | 0 |
| Step Delay | 0 | 0 | 0 |
| Aquisition Mode | None | None | Cont. |

Cooling

| Parameter | Value |
|------------------------------|-----------|
| Cycles | 1 |
| Type | Regular |
| Temp. Targets | Segment 1 |
| Target Temperature | 40 |
| Incubation time (h:min:s) | 30 |
| Temp. Transition Rate (°C/s) | 20 |
| Secondary Target Temp. | 0 |
| Step Size | 0 |
| Step Delay | 0 |
| Aquisition Mode | None |

| | |
|--------------------------------------|---|
| Preparation of the master mix | Depending on the total number of reactions place LightCycler™ capillaries in precooled centrifuge adaptors. It is recommended to use electronic pipettors with high quality tips (low volume retention). Prepare a master mix by multiplying the amount in the “Volume” column by the number of reactions to be analyzed, plus five additional reactions (Standard). |
|--------------------------------------|---|

| Step | Action | | | | | | | | | | |
|---|---|-----------|------|------------------------------|-----------|--------------------------------------|-----------|---|-----------|---------------------|-----------------------------|
| 1 | Prepare a fresh dilution series of the standard using the standard stabilizer solution 1:10 = 4000 copies/ μ l 1:100 = 400 copies/ μ l 1:1000 = 40 copies/ μ l | | | | | | | | | | |
| 2 | In a 1.5 ml light protected reaction tube on ice, add the following components in the order mentioned below: <table border="1" data-bbox="188 981 700 1176"> <thead> <tr> <th>Component</th> <th>Vol.</th> </tr> </thead> <tbody> <tr> <td>H₂O (white cap)</td> <td>6 μl</td> </tr> <tr> <td>LightCycler™ Primer Set (yellow cap)</td> <td>2 μl</td> </tr> <tr> <td>LightCycler™ FastStart DNA Master Sybr®Green I (premixed)</td> <td>2 μl</td> </tr> <tr> <td>Total Volume</td> <td>10 μl</td> </tr> </tbody> </table> | Component | Vol. | H ₂ O (white cap) | 6 μ l | LightCycler™ Primer Set (yellow cap) | 2 μ l | LightCycler™ FastStart DNA Master Sybr®Green I (premixed) | 2 μ l | Total Volume | 10 μl |
| Component | Vol. | | | | | | | | | | |
| H ₂ O (white cap) | 6 μ l | | | | | | | | | | |
| LightCycler™ Primer Set (yellow cap) | 2 μ l | | | | | | | | | | |
| LightCycler™ FastStart DNA Master Sybr®Green I (premixed) | 2 μ l | | | | | | | | | | |
| Total Volume | 10 μl | | | | | | | | | | |
| 3 | <ul style="list-style-type: none"> Pipet 10 μl PCR mix into the precooled LightCycler™ capillary Add 10 μl of cDNA template | | | | | | | | | | |
| 4 | <ul style="list-style-type: none"> Pipet 10 μl of PCR mix into 4 precooled LightCycler™ capillaries Add 10 μl of undiluted and of the freshly diluted standards into each capillary | | | | | | | | | | |
| 5 | Seal each capillary with a stopper and place the adaptors, containing the capillary, into a benchtop microcentrifuge. Centrifuge at 2000 rpm for 30 s. | | | | | | | | | | |
| 6 | Place capillaries in the rotor of the LightCycler™ Instrument. | | | | | | | | | | |
| 7 | Cycle the samples as described above | | | | | | | | | | |

| Typical results | |
|---------------------|---|
| Introduction | The analysis of the obtained data is divided into two parts: <ul style="list-style-type: none"> Part 1: Use of the quantification program, followed by Part 2: Specificity control of the amplification reaction by using the melting curve program |

| | |
|-------------------------------|---|
| Quantification program | The attached amplification curves in the QC sheet were obtained by performing the described procedure with the enclosed standards and control cDNA. The fluorescence values versus cycle number are displayed. The enclosed control cDNA contains approximately 200 copies per μ l of Bcl-10 specific cDNA |
| Melting curve program | Assess the specificity of the amplified PCR product by performing a melting curve analysis. The resulting melting curves allow discrimination between specific and unspecific product The attached melting curves in the QC sheet display the amplification of the control cDNA. As a control for the specificity, 5ng of human genomic DNA was amplified in this experiment. |



**GmbH
Heidelberg**

c/o Im Neuenheimer Feld 305

69120 Heidelberg

phone: +49-6221-418 604

Technical Hotline: +49-6221-418 617

Fax: +49-6221-418 608

www.search-lc.com info@search-lc.com

