LightCyclerTM -Primer Set

Ready-to-use amplification primer mix for RT-PCR using the LightCycler™ Instrument

Human HSP-70 (heat shock protein 70)

Kit for 96 reactions

Lot # 130302 Exp. 13.03.2003

Note: After Thawing keep on ice!

Store the kit at -20°C

1.Kit Cont	ents					
caution	After Thawing keep on ice!		Sample material	cDNA reverse transcribed from human RNA		
Kit contents	2 3	Label HSP-70 Primer mix Yellow cap Standard Red cap Standard Stabilizer Green cap Control cDNA Blue cap	Content and use 200 µl ready-to-use primer mix for target specific amplification using the LightCycler TM FastStart Master Sybr Green 1 contains optimal MgCl ₂ concentration and amplification primer pair 60 µl amplification standard for approximately 30000 copies/µl of HSP-70 cDNA 300 µl Solution for dilution of standard 50 µl contains a cDNA mix from several human hematopoietic cell lines	Sample Preparation !	Reliable and reproducible results are achieved with 1µg total RNA isolated with the HighPure total RNA Isolation Kit (Roche) reverse transcribed with the 1 st Strand cDNA Synthesis Kit (AMV) (Roche). The resulting cDNA has to be diluted to a final volume of 200-500 µl with PCR-grade water Quantitative evaluation of gene expression	
Additional equipment and reagents required	5 H2O, sterile, PCR grade White cap to adjust the final reaction volume White cap 1st Strand cDNA Synthesis Kit for RT-PCR (Roche Cat. # 1 483 188) LightCycler TM FastStart Master SybrGreen I (Roche Cat. # 3 003 230) LightCycler TM Instrument (Roche Cat. # 2 011 468) LightCycler TM Primer Set Housekeeping genes (Search GmbH)		Assay time Number of tests	in human cells and tissue Set up the PCR amplification 15 min LightCycler™ PCR run 50 min The Kit is designed for 96 Reactions		
2. Introduction The LightCycler TM -Primer Set allows to perform quantitative RT-PCR using the LightCycler TM instrument. An optimized primer pair has been selected for specific amplification of targets. The amplicon is detected by fluorescence using the double-stranded DNA binding dye Sybr®Green I.			ler TM instrument. An optimized I for specific amplification of cted by fluorescence using the	Quality Control Kit storage/ stability	The LightCycler TM -Primer Set is tested using the LightCycler TM FastStart Master Sybr®Green I according to the protocol described below. The unopened kit is stable at –20°C 12 month from date of manufacture	
				Specificity	The LightCycler TM -Primer Set "HSP-70" is specific for the sequence of human HSP-70 and does not detect genomic HSP-70 specific sequences if used as directed.	

3. Procedure			
Introduction	A fragment of the human HSP-70 cDNA sequence is amplified and monitored with the dsDNA specific Sybr®Green I dye		
Additional reagents required	LightCycler™ FastStart Master Sybr®Green I (Cat.# 3 003 230)		
Thawing the solutions	Thaw the following reagents, mix gently, and store on ice:		
	From the Thaw the		
	LightCycler TM FastStart vial 1a/b Master Sybr [®] Green I		
	LightCycler TM Primer Set all tubes		
	It is recommended to define the experimental protocol before preparing the solutions		
Experimental Protocol	 The described protocol consists of four programs. Program 1: Denaturation of the template and activation of the polymerase Program 2: Amplification of the target Program 3: Melting curve analysis for product control Program 4: Cooling the rotor 		
	and thermal chamber		

Denaturation

Parameter	Value
Cycles	1
Туре	Regular
Temp. Targets	Segment 1
Target Temperature	95
Incubation time (h:min:s)	10:00
Temp. Transition Rate (°C/s)	20
Secondary Target Temp.	0
Step Size	0
Step Delay	0
Aquisition Mode	None

Amplification

Parameter		Value	
Cycles		35	
Type	Q	uantificatio	on
Temp. Targets	Seg.1	Seg.2	Seg.3
Target Temperature	95	68	72
Incubation time (h:min:s)	10	10	16
Temp. Transition Rate (°C/s)	20	20	20
Secondary Target Temp.	0	58	0
Step Size	0	0.5	0
Step Delay	0	1	0
Aquisition Mode	None	None	Single
Gains		F1 = 5	

Melting Curve Analysis

Parameter		Value	
Cycles		1	
Туре	N.	lelting Cur	ve
Temp. Targets	Seg.1	Seg. 2	Seg.3
Target Temperature	95	58	95
Incubation time (h:min:s)	0	10	0
Temp. Transition Rate (°C/s)	20	20	0.1
Secondary Target Temp.	0	0	0
Step Size	0	0	0
Step Delay	0	0	0
Aquisition Mode	None	None	Cont.

Cooling

Parameter	Value
Cycles	1
Туре	Regular
Temp. Targets	Segment 1
Target Temperature	40
Incubation time (h:min:s)	30
Temp. Transition Rate (°C/s)	20
Secondary Target Temp.	0
Step Size	0
Step Delay	0
Aquisition Mode	None

D 0	TS 11 1 1 1 C
Preparation of	Depending on the total number of
the master	reactions place LightCycler TM
mix	capillaries in precooled centrifuge
	adaptors.
	It is recommended to use
	electronic pipettors with high
	quality tips (low volume
	retention). Prepare a master mix by
	multiplying the amount in the
	"Volume" column by the number
	of reactions to be analyzed, plus
	five additional reactions
	(Standard).

Step	Action		
1	Prepare a fresh dilution series of the standard		
	using the standard stabilizer solution		
	1:10 = 3000 copies/µl		
	1:100 = 300 copies/μl		
	1:1000 = 30 copies/μl		
2	In a 1.5 ml light protected reaction tube		
	ice, add the following components in th	e	
	order mentioned below:		
	Component	Vol.	
	H ₂ O (white cap)	6 µl	
	LightCycler TM Primer Set (yellow cap)	2 μl	
	LightCycler TM FastStart DNA Master	2 µl	
	Sybr®Green I (premixed)		
	Total Volume	10 µl	
3	• Pipet 10 µl PCR mix into the preco	oled	
	LightCycler™ capillary		
	 Add 10 μl of cDNA template 		
4	• Pipet 10 µl of PCR mix into 4 prece	ooled	
	LightCycler™ capillaries		
	• Add 10 µl of undiluted and of the f		
	diluted standards into each capillar	y	
5	Seal each capillary with a stopper and place		
	the adaptors, containing the capillary, ir	ito a	
	benchtop microcentrifuge. Centrifuge at	2000	
	rpm for 30 s.		
6	Place capillaries in the rotor of the		
	LightCycler TM Instrument.		
7	Cycle the samples as described above		
	-		



Typical results	
Introduction	The analysis of the obtained data is divided into two parts: • Part 1: Use of the quantification program, followed by • Part 2: Specificity control of the amplification reaction by using the melting curve program

Quantification program	The attached amplification curves in the QC sheet were obtained by performing the described procedure with the enclosed standards and control cDNA. The fluorescence values versus cycle number are displayed. The enclosed control cDNA contains approximately 3500 copies per µl of HSP-70 specific cDNA
Melting curve program	Assess the specificity of the amplified PCR product by performing a melting curve analysis. The resulting melting curves allow discrimination between specific and unspecific product The attached melting curves in the QC sheet display the amplification of the control cDNA. As a control for the specificity, 5ng of human genomic DNA was amplified in this experiment.

c/o Im Neuenheimer Feld 305 69120 Heidelberg

phone: +49-6221-418 604
Technical Hotline: +49-6221-418 617
Fax: +49-6221-418 608
www.search-lc.com info@search-lc.com

